

**DEPARTMENT OF ENVIRONMENTAL QUALITY - WATER DIVISION
EQUIPMENT TEMPERATURE LOG/THERMOMETER VERIFICATION CHECK SHEET**

1/08

FACILITY NAME:						VPDES NO:		VA00		DATE:			
EQUIPMENT	RANGE	IN RANGE		INSPECT READING °C	CHECK & LOG DAILY		CORRECT INCREMENT		ANNUAL THERMOMETER VERIFICATION				
		Y	N		Y	N	Y	N	Is the NIST / NIST-Traceable Reference Thermometer within the manufacturer's expiration date or recertified yearly?			Y/N	
									DATE CHECKED	MARKED		CORR FACTOR °C	INSPECT TEMP °C
										Y	N		
SAMPLE REFRIGER.	1-6°C												
AUTO SAMPLER	1-6° C												
BOD INCUBATOR	20 ± 1° C												
SOLIDS DRYING OVEN	103-105° C												
WATER BATH	44.5 ± .2° C												
INCUBATOR	35± .5° C												
AUTOClave	121° C IN 30 MIN												
HOT AIR STERILIZING	170 ± 10° C												
O & G WATER BATH	70± 2° C												
REAGENT REFRIGER.	1-6° C												
pH METER	± 1° C												
DO METER	± 1° C												
THERMOMETER-OUTFALL	± 1° C												
Hg WATER BATH	95 °C												

PROBLEMS:

templog